

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN121718\
 Data File : VN052933.D
 Acq On : 17 Dec 2018 17:51
 Operator : MD\SY
 Sample : PB115729ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115729ZHE#03

Quant Time: Dec 18 06:09:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1327406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2063600	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1828415	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	720459	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	672070	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
35) Dibromofluoromethane	7.59	113	466735	40.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.04%	
50) Toluene-d8	10.09	98	2509870	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
62) 4-Bromofluorobenzene	12.40	95	816237	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
Target Compounds						
16) Acetone	3.82	43	23662	7.465	ug/l	98
20) Methylene Chloride	4.55	84	22834	1.601	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN121718\
Data File : VN052933.D
Acq On : 17 Dec 2018 17:51
Operator : MD\SY
Sample : PB115729ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115729ZHE#03

Quant Time: Dec 18 06:09:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

