

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN121718\
 Data File : VN052935.D
 Acq On : 17 Dec 2018 18:45
 Operator : MD\SY
 Sample : PB115729ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115729ZHE#05

Quant Time: Dec 18 06:11:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1338343	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2088577	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1797468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	701292	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	694197	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
35) Dibromofluoromethane	7.59	113	460467	39.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.00%	
50) Toluene-d8	10.09	98	2508986	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
62) 4-Bromofluorobenzene	12.40	95	787270	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
Target Compounds						
16) Acetone	3.82	43	22180	6.940	ug/l	98
20) Methylene Chloride	4.56	84	23458	1.631	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN121718\
Data File : VN052935.D
Acq On : 17 Dec 2018 18:45
Operator : MD\SY
Sample : PB115729ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115729ZHE#05

Quant Time: Dec 18 06:11:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

