

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121719\
 Data File : VN059620.D
 Acq On : 17 Dec 2019 13:16
 Operator : JC/SP
 Sample : VN1217MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1217MBL01

Quant Time: Dec 18 05:58:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	222277	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	353757	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	310680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	115137	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	155611	48.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.22%	
35) Dibromofluoromethane	7.56	113	113246	51.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.32%	
50) Toluene-d8	10.08	98	436737	49.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.68%	
62) 4-Bromofluorobenzene	12.40	95	139335	43.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.14%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121719\
Data File : VN059620.D
Acq On : 17 Dec 2019 13:16
Operator : JC/SP
Sample : VN1217MBL01
Misc : 5.00µ/10mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1217MBL01

Quant Time: Dec 18 05:58:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

