

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121719\
 Data File : VN059625.D
 Acq On : 17 Dec 2019 15:30
 Operator : JC/SP
 Sample : PB125506TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB125506TB

Quant Time: Dec 18 06:06:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	223604	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	361658	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	325836	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	115588	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	163392	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
35) Dibromofluoromethane	7.57	113	113702	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
50) Toluene-d8	10.08	98	455641	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
62) 4-Bromofluorobenzene	12.40	95	144406	44.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.34%	
Target Compounds						
20) Methylene Chloride	4.52	84	4073	1.460	ug/l	85
43) Isopropyl Acetate	8.14	43	74619	10.949	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121719\
Data File : VN059625.D
Acq On : 17 Dec 2019 15:30
Operator : JC/SP
Sample : PB125506TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB125506TB

Quant Time: Dec 18 06:06:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

