

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121719\
 Data File : VN059627.D
 Acq On : 17 Dec 2019 16:14
 Operator : JC/SP
 Sample : K6294-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Dec 18 06:10:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	233622	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	376364	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	330097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	126311	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	169516	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
35) Dibromofluoromethane	7.57	113	118532	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
50) Toluene-d8	10.08	98	460856	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
62) 4-Bromofluorobenzene	12.40	95	148645	43.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.38%	
Target Compounds						
16) Acetone	3.79	43	16153	8.003	ug/l	100
20) Methylene Chloride	4.52	84	91551	31.407	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121719\
Data File : VN059627.D
Acq On : 17 Dec 2019 16:14
Operator : JC/SP
Sample : K6294-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Dec 18 06:10:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

