

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121719\
 Data File : VN059634.D
 Acq On : 17 Dec 2019 18:48
 Operator : JC/SP
 Sample : K6330-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11977

Quant Time: Dec 18 06:24:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	227477	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	368145	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	330389	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	123985	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	170348	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
35) Dibromofluoromethane	7.57	113	119766	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
50) Toluene-d8	10.08	98	467493	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
62) 4-Bromofluorobenzene	12.40	95	147753	44.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.80%	
Target Compounds						
16) Acetone	3.80	43	14495	7.376	ug/l	97
18) Methyl Acetate	4.30	43	6718	1.631	ug/l #	68
20) Methylene Chloride	4.53	84	8549	3.012	ug/l	96
43) Isopropyl Acetate	8.15	43	70078	10.101	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121719\
Data File : VN059634.D
Acq On : 17 Dec 2019 18:48
Operator : JC/SP
Sample : K6330-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11977

Quant Time: Dec 18 06:24:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

