

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070156.D
 Acq On : 17 Dec 2021 11:32
 Operator : JC/MD
 Sample : VN1217WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1217WBL01

Quant Time: Dec 18 05:29:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Dec 14 00:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.659	128	216539	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	1238840	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.747	117	1070319	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	601755	31.358	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.533%
60) 4-Bromofluorobenzene	12.733	95	441955	25.094	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.633%
63) Toluene-d8	10.443	98	1520261	30.736	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.467%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070156.D
Acq On : 17 Dec 2021 11:32
Operator : JC/MD
Sample : VN1217WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1217WBL01

Quant Time: Dec 18 05:29:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Dec 14 00:59:04 2021
Response via : Initial Calibration

