

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070175.D
 Acq On : 17 Dec 2021 20:00
 Operator : JC/MD
 Sample : M5039-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-31

Quant Time: Dec 18 05:48:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	918898	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1551443	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1366929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	474705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	703329	49.573	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.140%		
35) Dibromofluoromethane	8.024	113	454586	46.778	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.560%		
50) Toluene-d8	10.443	98	1732450	44.398	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	88.800%		
62) 4-Bromofluorobenzene	12.731	95	569345	40.275	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	80.560%		
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	26606	3.117	ug/l	97
40) Benzene	8.461	78	17247	0.382	ug/l	95
52) Toluene	10.508	92	30913	1.122	ug/l	100
67) Ethyl Benzene	11.846	91	76135	1.511	ug/l	98
68) m/p-Xylenes	11.956	106	86701	4.727	ug/l	90
69) o-Xylene	12.283	106	21416	1.159	ug/l	94
78) n-propylbenzene	12.921	91	30025	0.639	ug/l	97
80) 1,3,5-Trimethylbenzene	13.061	105	31109	0.915	ug/l	98
84) 1,2,4-Trimethylbenzene	13.369	105	144727	4.470	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070175.D
Acq On : 17 Dec 2021 20:00
Operator : JC/MD
Sample : M5039-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-31

Quant Time: Dec 18 05:48:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

