

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070181.D
 Acq On : 17 Dec 2021 23:25
 Operator : JC/MD
 Sample : VN1217WBL03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1217WBL03

Quant Time: Dec 18 06:01:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	985480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1676175	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1477382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	502221	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	812009	53.366	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.740%
35) Dibromofluoromethane	8.021	113	534584	50.917	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.840%
50) Toluene-d8	10.443	98	2039769	48.384	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.760%
62) 4-Bromofluorobenzene	12.733	95	666105	43.614	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070181.D
Acq On : 17 Dec 2021 23:25
Operator : JC/MD
Sample : VN1217WBL03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1217WBL03

Quant Time: Dec 18 06:01:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

