

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070184.D
 Acq On : 18 Dec 2021 00:42
 Operator : JC/MD
 Sample : M5128-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Dec 18 06:02:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1082977	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2036632	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2360605	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	966056	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	839046	50.179	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.360%
35) Dibromofluoromethane	8.021	113	596177	46.733	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.460%
50) Toluene-d8	10.443	98	2686761	52.452	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.900%
62) 4-Bromofluorobenzene	12.731	95	1183569	63.779	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	127.560%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	39447	3.922	ug/l	94
93) 1,2,4-Trichlorobenzene	15.265	180	10455	3.500	ug/l	98
95) Naphthalene	15.507	128	63420	5.221	ug/l	98
96) 1,2,3-Trichlorobenzene	15.700	180	12706	3.218	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070184.D
Acq On : 18 Dec 2021 00:42
Operator : JC/MD
Sample : M5128-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Dec 18 06:02:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

