

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070185.D
 Acq On : 18 Dec 2021 1:07
 Operator : JC/MD
 Sample : M5128-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Dec 18 06:02:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1155666	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2189617	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2595010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1090585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	906498	50.803	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.600%
35) Dibromofluoromethane	8.024	113	640690	46.714	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.420%
50) Toluene-d8	10.443	98	2889746	52.473	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.940%
62) 4-Bromofluorobenzene	12.731	95	1309043	65.612	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	131.220%
Target Compounds						
95) Naphthalene	15.504	128	32331	4.543	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070185.D
Acq On : 18 Dec 2021 1:07
Operator : JC/MD
Sample : M5128-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Dec 18 06:02:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

