

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070187.D
 Acq On : 18 Dec 2021 1:59
 Operator : JC/MD
 Sample : M5128-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Dec 18 06:02:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1133029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2161021	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2543396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1065415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	898098	51.338	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.680%
35) Dibromofluoromethane	8.021	113	633168	46.776	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.560%
50) Toluene-d8	10.443	98	2840886	52.268	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.540%
62) 4-Bromofluorobenzene	12.731	95	1289817	65.504	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	131.000%
Target Compounds						
						Qvalue
16) Acetone	4.301	43	27513	2.614	ug/l	100
95) Naphthalene	15.507	128	15340	4.256	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070187.D
Acq On : 18 Dec 2021 1:59
Operator : JC/MD
Sample : M5128-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Dec 18 06:02:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

