

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070190.D
 Acq On : 18 Dec 2021 3:15
 Operator : JC/MD
 Sample : M5132-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: Dec 19 07:22:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1132033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2156954	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2562247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1072749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	889876	50.912	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.820%
35) Dibromofluoromethane	8.021	113	634735	46.980	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.960%
50) Toluene-d8	10.443	98	2857023	52.664	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.320%
62) 4-Bromofluorobenzene	12.731	95	1287426	65.506	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	131.020%
Target Compounds						
16) Acetone	4.299	43	55462	5.275	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070190.D
Acq On : 18 Dec 2021 3:15
Operator : JC/MD
Sample : M5132-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: Dec 19 07:22:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

