

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070192.D
 Acq On : 18 Dec 2021 4:07
 Operator : JC/MD
 Sample : M5132-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Dec 19 07:22:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1140326	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2169273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2588822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1054128	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	910312	51.703	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.400%	
35) Dibromofluoromethane	8.024	113	637046	46.883	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 93.760%	
50) Toluene-d8	10.441	98	2891498	52.997	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.000%	
62) 4-Bromofluorobenzene	12.731	95	1288535	65.190	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 130.380%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.446	62	99462	6.492	ug/l	99
12) 1,1-Dichloroethene	4.183	96	336611	30.990	ug/l	87
16) Acetone	4.293	43	36938	3.488	ug/l	98
24) 1,1-Dichloroethane	6.388	63	139577	5.632	ug/l	98
27) cis-1,2-Dichloroethene	7.324	96	94679	6.642	ug/l	95
40) Benzene	8.461	78	24480	0.388	ug/l	99
42) 1,2-Dichloroethane	8.534	62	44756	1.756	ug/l	97
44) Trichloroethene	9.220	130	6579	0.437	ug/l	83
49) 1,4-Dioxane	9.572	88	148389	389.607	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070192.D
Acq On : 18 Dec 2021 4:07
Operator : JC/MD
Sample : M5132-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Dec 19 07:22:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

