

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070193.D
 Acq On : 18 Dec 2021 4:32
 Operator : JC/MD
 Sample : M5132-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3

Quant Time: Dec 19 07:23:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1175413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2219250	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2615309	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1073905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	933946	51.462	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.920%	
35) Dibromofluoromethane	8.024	113	655692	47.169	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 94.340%	
50) Toluene-d8	10.443	98	2938531	52.646	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.300%	
62) 4-Bromofluorobenzene	12.731	95	1295776	64.080	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 128.160%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.181	96	67363	6.017	ug/l	90
16) Acetone	4.301	43	26845	2.459	ug/l #	88
24) 1,1-Dichloroethane	6.388	63	36229	1.418	ug/l	97
27) cis-1,2-Dichloroethene	7.329	96	23339	1.588	ug/l	96
95) Naphthalene	15.512	128	7756	4.121	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070193.D
Acq On : 18 Dec 2021 4:32
Operator : JC/MD
Sample : M5132-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3

Quant Time: Dec 19 07:23:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

