

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070194.D
 Acq On : 18 Dec 2021 4:58
 Operator : JC/MD
 Sample : M5132-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Dec 19 07:23:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1125241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2161379	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2582089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1064041	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	899768	51.789	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.580%
35) Dibromofluoromethane	8.021	113	637377	47.079	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.160%
50) Toluene-d8	10.443	98	2892138	53.202	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.400%
62) 4-Bromofluorobenzene	12.733	95	1290405	65.523	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	131.040%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	4.181	96	5620	0.524	ug/l #	93
27) cis-1,2-Dichloroethene	7.327	96	6307	0.448	ug/l	97
49) 1,4-Dioxane	9.574	88	4688	12.354	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070194.D
Acq On : 18 Dec 2021 4:58
Operator : JC/MD
Sample : M5132-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Dec 19 07:23:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

