

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070195.D
 Acq On : 18 Dec 2021 5:23
 Operator : JC/MD
 Sample : M5132-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP-1

Quant Time: Dec 19 07:23:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1090965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2082310	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2454551	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	972238	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	868226	51.544	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.080%
35) Dibromofluoromethane	8.024	113	608615	46.662	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.320%
50) Toluene-d8	10.443	98	2785405	53.184	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.360%
62) 4-Bromofluorobenzene	12.731	95	1201031	63.301	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	126.600%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	4.181	96	5526	0.532	ug/l	91
16) Acetone	4.293	43	21504	2.122	ug/l	98
27) cis-1,2-Dichloroethene	7.327	96	5633	0.413	ug/l	94
49) 1,4-Dioxane	9.577	88	3859	10.555	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070195.D
Acq On : 18 Dec 2021 5:23
Operator : JC/MD
Sample : M5132-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP-1

Quant Time: Dec 19 07:23:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

