

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070199.D
 Acq On : 18 Dec 2021 7:06
 Operator : JC/MD
 Sample : M5115-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/20/2021
 Supervised By : Mahesh Dadoda 12/23/2021

Quant Time: Dec 19 07:25:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1087661	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2064019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2448251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1030787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	855596	50.948	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.900%			
35) Dibromofluoromethane	8.024	113	609791	47.166	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 94.340%			
50) Toluene-d8	10.441	98	2752915	53.030	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.060%			
62) 4-Bromofluorobenzene	12.731	95	1247545	66.335	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 132.660%			
Target Compounds					Qvalue	
16) Acetone	4.301	43	38074	3.769	ug/l	96
19) Methyl tert-butyl Ether	5.629	73	51984m	1.290	ug/l	
95) Naphthalene	15.512	128	5861	4.092	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070199.D
Acq On : 18 Dec 2021 7:06
Operator : JC/MD
Sample : M5115-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/20/2021
Supervised By :Mahesh Dadoda 12/23/2021

Quant Time: Dec 19 07:25:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

