

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070201.D
 Acq On : 18 Dec 2021 7:57
 Operator : JC/MD
 Sample : M5115-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7

Quant Time: Dec 19 07:25:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1120148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2141908	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2553458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1148363	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	889549	51.434	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	102.860%	
35) Dibromofluoromethane	8.021	113	628878	46.874	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	93.740%	
50) Toluene-d8	10.443	98	2843577	52.784	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	105.560%	
62) 4-Bromofluorobenzene	12.731	95	1334109	68.358	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	136.720%#	
Target Compounds						
						Qvalue
16) Acetone	4.299	43	13216	1.270	ug/l	97
73) Isopropylbenzene	12.578	105	82807	0.810	ug/l	98
85) sec-Butylbenzene	13.501	105	83837	0.849	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070201.D
Acq On : 18 Dec 2021 7:57
Operator : JC/MD
Sample : M5115-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7

Quant Time: Dec 19 07:25:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

