

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070203.D
 Acq On : 18 Dec 2021 8:48
 Operator : JC/MD
 Sample : M5115-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2021-12-15

Quant Time: Dec 19 07:26:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1144240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2181814	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2581918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1079073	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	901227	51.012	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.020%	
35) Dibromofluoromethane	8.024	113	631939	46.240	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 92.480%	
50) Toluene-d8	10.443	98	2898507	52.820	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.640%	
62) 4-Bromofluorobenzene	12.731	95	1300629	65.424	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 130.840%	
Target Compounds						
16) Acetone	4.304	43	8095	0.762	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070203.D
Acq On : 18 Dec 2021 8:48
Operator : JC/MD
Sample : M5115-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2021-12-15

Quant Time: Dec 19 07:26:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

