

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070180.D
 Acq On : 17 Dec 2021 22:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 18 06:00:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.583	0.492	15.6	67	0.00
3 P	Chloromethane	0.731	0.610	16.6	72	0.00
4 C	Vinyl Chloride	0.672	0.575	14.4#	72	0.00
5 T	Bromomethane	0.365	0.314	14.0	72	0.00
6 T	Chloroethane	0.412	0.358	13.1	74	0.00
7 T	Trichlorofluoromethane	0.882	0.779	11.7	75	0.00
8 T	Diethyl Ether	0.340	0.359	-5.6	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.488	1.2	83	0.00
10 T	Methyl Iodide	0.675	0.645	4.4	76	0.00
11 T	Tert butyl alcohol	0.146	0.155	-6.2	88	0.00
12 CM	1,1-Dichloroethene	0.476	0.462	2.9#	82	0.00
13 T	Acrolein	0.124	0.110	11.3	79	0.00
14 T	Allyl chloride	1.073	1.093	-1.9	86	0.00
15 T	Acrylonitrile	0.393	0.415	-5.6	87	0.00
16 T	Acetone	0.464	0.393	15.3	70	0.00
17 T	Carbon Disulfide	1.230	1.015	17.5	68	0.00
18 T	Methyl Acetate	1.402	1.437	-2.5	88	0.00
19 T	Methyl tert-butyl Ether	1.852	2.014	-8.7	91	0.00
20 T	Methylene Chloride	0.617	0.594	3.7	86	0.00
21 T	trans-1,2-Dichloroethene	0.517	0.486	6.0	79	0.00
22 T	Diisopropyl ether	2.098	2.307	-10.0	90	0.00
23 T	Vinyl Acetate	1.707	1.855	-8.7	87	0.00
24 P	1,1-Dichloroethane	1.087	1.128	-3.8	87	0.00
25 T	2-Butanone	0.631	0.624	1.1	81	0.00
26 T	2,2-Dichloropropane	0.845	0.784	7.2	78	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.628	-0.5	84	0.00
28 T	Bromochloromethane	0.565	0.529	6.4	79	0.00
29 T	Tetrahydrofuran	0.383	0.402	-5.0	85	0.00
30 C	Chloroform	1.097	1.139	-3.8#	87	0.00
31 T	Cyclohexane	1.074	0.980	8.8	78	0.00
32 T	1,1,1-Trichloroethane	0.935	0.959	-2.6	84	0.00
33 S	1,2-Dichloroethane-d4	0.772	0.848	-9.8	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.313	0.345	-10.2	93	0.00
36 T	1,1-Dichloropropene	0.476	0.467	1.9	81	0.00
37 T	Ethyl Acetate	0.708	0.716	-1.1	83	0.00
38 T	Carbon Tetrachloride	0.467	0.457	2.1	80	0.00
39 T	Methylcyclohexane	0.566	0.542	4.2	77	0.00
40 TM	Benzene	1.455	1.457	-0.1	83	0.00
41 T	Methacrylonitrile	0.333	0.378	-13.5	91	0.00
42 TM	1,2-Dichloroethane	0.587	0.600	-2.2	86	0.00
43 T	Isopropyl Acetate	1.048	1.127	-7.5	88	0.00
44 TM	Trichloroethene	0.347	0.338	2.6	80	0.00
45 C	1,2-Dichloropropane	0.391	0.410	-4.9#	88	0.00
46 T	Dibromomethane	0.256	0.261	-2.0	84	0.00
47 T	Bromodichloromethane	0.485	0.526	-8.5	87	0.00
48 T	Methyl methacrylate	0.489	0.527	-7.8	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
 Data File : VN070180.D
 Acq On : 17 Dec 2021 22:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 18 06:00:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	85	0.00
50 S	Toluene-d8	1.258	1.348	-7.2	90	0.00
51 T	4-Methyl-2-Pentanone	0.680	0.759	-11.6	88	0.00
52 CM	Toluene	0.888	0.902	-1.6#	83	0.00
53 T	t-1,3-Dichloropropene	0.510	0.555	-8.8	84	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.596	-8.8	84	0.00
55 T	1,1,2-Trichloroethane	0.358	0.378	-5.6	87	0.00
56 T	Ethyl methacrylate	0.608	0.667	-9.7	87	0.00
57 T	1,3-Dichloropropane	0.631	0.666	-5.5	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.154	0.204	-32.5#	104	0.00
59 T	2-Hexanone	0.514	0.558	-8.6	85	0.00
60 T	Dibromochloromethane	0.340	0.369	-8.5	85	0.00
61 T	1,2-Dibromoethane	0.371	0.384	-3.5	84	0.00
62 S	4-Bromofluorobenzene	0.456	0.516	-13.2	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.339	0.331	2.4	81	0.00
65 PM	Chlorobenzene	1.014	1.029	-1.5	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.378	-5.9	84	0.00
67 C	Ethyl Benzene	1.843	1.942	-5.4#	84	0.00
68 T	m/p-Xylenes	0.671	0.707	-5.4	83	0.00
69 T	o-Xylene	0.676	0.720	-6.5	85	0.00
70 T	Styrene	1.072	1.181	-10.2	85	0.00
71 P	Bromoform	0.264	0.298	-12.9	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	4.449	4.430	0.4	85	0.00
74 T	N-amyl acetate	1.967	2.154	-9.5	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.538	1.549	-0.7	88	0.00
76 T	1,2,3-Trichloropropane	1.434	1.354	5.6	80	0.00
77 T	Bromobenzene	1.045	1.018	2.6	85	0.00
78 T	n-propylbenzene	4.950	5.006	-1.1	84	0.00
79 T	2-Chlorotoluene	3.128	3.113	0.5	86	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.626	-1.3	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.396	0.408	-3.0	81	0.00
82 T	4-Chlorotoluene	2.942	3.005	-2.1	86	0.00
83 T	tert-Butylbenzene	3.171	3.205	-1.1	85	0.00
84 T	1,2,4-Trimethylbenzene	3.410	3.574	-4.8	86	0.00
85 T	sec-Butylbenzene	4.301	4.432	-3.0	85	0.00
86 T	p-Isopropyltoluene	3.362	3.497	-4.0	85	0.00
87 T	1,3-Dichlorobenzene	1.717	1.709	0.5	84	0.00
88 T	1,4-Dichlorobenzene	1.687	1.675	0.7	85	0.00
89 T	n-Butylbenzene	2.727	2.874	-5.4	85	0.00
90 T	Hexachloroethane	0.590	0.587	0.5	80	0.00
91 T	1,2-Dichlorobenzene	1.673	1.703	-1.8	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.278	0.297	-6.8	88	0.00
93 T	1,2,4-Trichlorobenzene	0.760	0.825	-8.6	90	0.00
94 T	Hexachlorobutadiene	0.443	0.471	-6.3	88	0.00
95 T	Naphthalene	2.199	2.571	-16.9	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121721\
Data File : VN070180.D
Acq On : 17 Dec 2021 22:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 18 06:00:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.838	-11.9	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6