

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121724\
 Data File : VN085226.D
 Acq On : 17 Dec 2024 15:56
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 18 00:27:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	142514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	233387	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	215690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	101155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	95279	46.935	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.880%		
35) Dibromofluoromethane	8.159	113	76698	48.732	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.460%		
50) Toluene-d8	10.565	98	284551	50.448	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.900%		
62) 4-Bromofluorobenzene	12.847	95	105002	49.780	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.560%		
Target Compounds						
					Qvalue	
10) Methyl Iodide	4.571	142	3723	1.736	ug/l #	92
67) Ethyl Benzene	11.953	91	1666	0.210	ug/l #	83
68) m/p-Xylenes	12.071	106	1731	0.573	ug/l	87
78) n-propylbenzene	13.035	91	2013	0.245	ug/l	80
84) 1,2,4-Trimethylbenzene	13.482	105	9157	1.486	ug/l	98
95) Naphthalene	15.635	128	278307	49.554	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121724\
Data File : VN085226.D
Acq On : 17 Dec 2024 15:56
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 18 00:27:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

