

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121724\
 Data File : VN085229.D
 Acq On : 17 Dec 2024 17:09
 Operator : JC\MD
 Sample : P5291-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EQUIP-BLANK-20241213

Quant Time: Dec 18 00:28:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262628	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	225155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93992	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107339	49.329	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.660%
35) Dibromofluoromethane	8.165	113	86748	48.980	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.960%
50) Toluene-d8	10.565	98	306559	48.299	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.600%
62) 4-Bromofluorobenzene	12.847	95	105993	44.655	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.300%
Target Compounds						
16) Acetone	4.442	43	3552	5.605	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121724\
Data File : VN085229.D
Acq On : 17 Dec 2024 17:09
Operator : JC\MD
Sample : P5291-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EQUIP-BLANK-20241213

Quant Time: Dec 18 00:28:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

