

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN121818\
 Data File : VN052981.D
 Acq On : 19 Dec 2018 1:05
 Operator : MD\SY
 Sample : J6411-19DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-MW-FD-1218DL

Quant Time: Dec 19 12:26:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1320523	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2009362	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1840352	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	835863	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	642915	47.98	ug/l	0.00
Spiked Amount			Recovery	=	95.96%	
35) Dibromofluoromethane	7.59	113	509581	55.52	ug/l	0.00
Spiked Amount			Recovery	=	111.04%	
50) Toluene-d8	10.09	98	2469826	50.12	ug/l	0.00
Spiked Amount			Recovery	=	100.24%	
62) 4-Bromofluorobenzene	12.40	95	868539	50.21	ug/l	0.00
Spiked Amount			Recovery	=	100.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
39) Methylcyclohexane	9.08	83	7495	0.320	ug/l	99
67) Ethyl Benzene	11.51	91	29903	0.426	ug/l	97
68) m/p-Xylenes	11.63	106	39343	1.475	ug/l	95
73) Isopropylbenzene	12.25	105	1176378	17.988	ug/l	100
78) n-propylbenzene	12.59	91	2466017	33.108	ug/l	100
83) tert-Butylbenzene	12.99	119	113551	2.409	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	2325370	43.394	ug/l	100
85) sec-Butylbenzene	13.17	105	642462	10.075	ug/l #	74
86) p-Isopropyltoluene	13.29	119	56747	1.030	ug/l	99
89) n-Butylbenzene	13.62	91	297373	6.273	ug/l #	84
95) Naphthalene	15.13	128	6812	1.349	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN121818\
Data File : VN052981.D
Acq On : 19 Dec 2018 1:05
Operator : MD\SY
Sample : J6411-19DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EP1-MW-FD-1218DL

Quant Time: Dec 19 12:26:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

