

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN121818\
 Data File : VN052985.D
 Acq On : 19 Dec 2018 2:52
 Operator : MD\SY
 Sample : PB115784TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115784TB

Quant Time: Dec 19 05:45:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1477522	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2254940	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2019851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	814238	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	720741	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
35) Dibromofluoromethane	7.59	113	513501	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
50) Toluene-d8	10.09	98	2754164	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
62) 4-Bromofluorobenzene	12.40	95	907972	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
Target Compounds						
16) Acetone	3.82	43	23794	8.094	ug/l	91
18) Methyl Acetate	4.33	43	6264	1.452	ug/l #	86
20) Methylene Chloride	4.56	84	40982	2.556	ug/l	95
43) Isopropyl Acetate	8.17	43	78155	2.347	ug/l #	84
95) Naphthalene	15.13	128	22008	1.818	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN121818\
Data File : VN052985.D
Acq On : 19 Dec 2018 2:52
Operator : MD\SY
Sample : PB115784TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115784TB

Quant Time: Dec 19 05:45:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

