

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
 Data File : VN070211.D
 Acq On : 18 Dec 2021 14:19
 Operator : JC/MD
 Sample : M5044-02 10X
 Misc : 5.53g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DA-2-4.5

Quant Time: Dec 19 08:45:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1118607	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2159660	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2629021	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1097212	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	880893	51.00	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.00%	
35) Dibromofluoromethane	8.021	113	617989	45.68	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 91.36%	
50) Toluene-d8	10.440	98	2893625	53.27	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.54%	
62) 4-Bromofluorobenzene	12.731	95	1370433	69.64	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 139.28%#	
Target Compounds						
95) Naphthalene	15.507	128	39486	4.66	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
Data File : VN070211.D
Acq On : 18 Dec 2021 14:19
Operator : JC/MD
Sample : M5044-02 10X
Misc : 5.53g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DA-2-4.5

Quant Time: Dec 19 08:45:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

