

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
 Data File : VN070213.D
 Acq On : 18 Dec 2021 15:10
 Operator : JC/MD
 Sample : M5044-04 10X
 Misc : 5.64g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DA-4-4.5

Quant Time: Dec 19 08:46:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	1169614	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2253005	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2789322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1239856	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	928071	51.39	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.78%	
35) Dibromofluoromethane	8.021	113	648559	45.96	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 91.92%	
50) Toluene-d8	10.440	98	3082607	54.40	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 108.80%	
62) 4-Bromofluorobenzene	12.731	95	1487429	72.46	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 144.92%#	
Target Compounds						
39) Methylcyclohexane	9.459	83	36678m	1.44	ug/l	Qvalue
95) Naphthalene	15.507	128	18834	4.27	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
Data File : VN070213.D
Acq On : 18 Dec 2021 15:10
Operator : JC/MD
Sample : M5044-04 10X
Misc : 5.64g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DA-4-4.5

Quant Time: Dec 19 08:46:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

