

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
 Data File : VN070219.D
 Acq On : 18 Dec 2021 17:43
 Operator : JC/MD
 Sample : M5044-10 10X
 Misc : 5.62g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DA-10-4.5

Quant Time: Dec 19 08:48:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1177822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2251244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2768153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1169971	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	932781	51.29	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.58%			
35) Dibromofluoromethane	8.021	113	640283	45.41	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 90.82%			
50) Toluene-d8	10.440	98	3044727	53.77	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.54%			
62) 4-Bromofluorobenzene	12.731	95	1418133	69.13	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 138.26%#			
Target Compounds					Qvalue	
95) Naphthalene	15.504	128	13904	4.21	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
Data File : VN070219.D
Acq On : 18 Dec 2021 17:43
Operator : JC/MD
Sample : M5044-10 10X
Misc : 5.62g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DA-10-4.5

Quant Time: Dec 19 08:48:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

