

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
 Data File : VN070223.D
 Acq On : 18 Dec 2021 19:26
 Operator : JC/MD
 Sample : M5044-14 10X
 Misc : 6.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DA-14-4.5

Quant Time: Dec 19 08:50:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1186248	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2206664	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2715331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1312950	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	905502	49.44	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.88%		
35) Dibromofluoromethane	8.021	113	633355	45.82	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.64%		
50) Toluene-d8	10.440	98	3032888	54.65	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	109.30%		
62) 4-Bromofluorobenzene	12.731	95	1477818	73.50	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	147.00%#		
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.459	83	554173	22.18	ug/l #	78
67) Ethyl Benzene	11.843	91	4454828	44.52	ug/l	96
68) m/p-Xylenes	11.958	106	1350052	37.06	ug/l	92
69) o-Xylene	12.280	106	39544	1.08	ug/l	98
73) Isopropylbenzene	12.578	105	687547	5.89	ug/l	97
78) n-propylbenzene	12.918	91	2887822	22.22	ug/l	95
80) 1,3,5-Trimethylbenzene	13.058	105	1228712	13.07	ug/l	99
84) 1,2,4-Trimethylbenzene	13.366	105	8233125	91.95	ug/l	98
85) sec-Butylbenzene	13.500	105	301984	2.67	ug/l #	60
86) p-Isopropyltoluene	13.616	119	134715	1.53	ug/l	96
89) n-Butylbenzene	13.943	91	813913	11.37	ug/l #	72
95) Naphthalene	15.504	128	1545896	26.16	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
Data File : VN070223.D
Acq On : 18 Dec 2021 19:26
Operator : JC/MD
Sample : M5044-14 10X
Misc : 6.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DA-14-4.5

Quant Time: Dec 19 08:50:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

