

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
 Data File : VN070227.D
 Acq On : 18 Dec 2021 21:08
 Operator : JC/MD
 Sample : M5044-18ME
 Misc : 6.69g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DA-18-5.5ME

Quant Time: Dec 19 08:52:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1080756	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2087917	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2563931	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1105127	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	861359	51.62	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.24%
35) Dibromofluoromethane	8.021	113	590737	45.17	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.34%
50) Toluene-d8	10.440	98	2839761	54.08	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.16%
62) 4-Bromofluorobenzene	12.731	95	1347809	70.85	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	141.70%#
Target Compounds						
89) n-Butylbenzene	13.940	91	92478	1.53	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121821\
Data File : VN070227.D
Acq On : 18 Dec 2021 21:08
Operator : JC/MD
Sample : M5044-18ME
Misc : 6.69g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DA-18-5.5ME

Quant Time: Dec 19 08:52:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

