

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121824\
 Data File : VN085261.D
 Acq On : 18 Dec 2024 19:33
 Operator : JC\MD
 Sample : P5291-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EQUIP-BLANK-20241213

Quant Time: Dec 19 06:49:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 19 06:31:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	102235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	187663	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	161222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	62465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	91200	51.038	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.080%			
35) Dibromofluoromethane	8.171	113	65713	49.806	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.620%			
50) Toluene-d8	10.565	98	229435	47.578	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.160%			
62) 4-Bromofluorobenzene	12.847	95	72343	42.002	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 84.000%			
Target Compounds						
16) Acetone	4.453	43	3668	5.947	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121824\
Data File : VN085261.D
Acq On : 18 Dec 2024 19:33
Operator : JC\MD
Sample : P5291-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EQUIP-BLANK-20241213

Quant Time: Dec 19 06:49:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121824W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 19 06:31:00 2024
Response via : Initial Calibration

