

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121918\
 Data File : VN053005.D
 Acq On : 19 Dec 2018 12:36
 Operator : MD\SY
 Sample : J6386-05ME
 Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-SS005-0006-01ME

Manual Integrations
 APPROVED

MMDadoda
 12/20/2018 9:34:32 AM

Quant Time: Dec 20 07:00:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1148139	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1754833	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1594748	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	692140	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	547036	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
35) Dibromofluoromethane	7.59	113	419800	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
50) Toluene-d8	10.09	98	2153892	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	12.40	95	720242	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	3168m	1.387	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121918\
Data File : VN053005.D
Acq On : 19 Dec 2018 12:36
Operator : MD\SY
Sample : J6386-05ME
Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
P001-SS005-0006-01ME

Manual Integrations
APPROVED

MMDadoda
12/20/2018 9:34:32 AM

Quant Time: Dec 20 07:00:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

