

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121918\
 Data File : VN053017.D
 Acq On : 19 Dec 2018 17:56
 Operator : MD\SY
 Sample : J6475-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SLUDGE

Quant Time: Dec 20 06:41:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 18 13:03:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1272093	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1965351	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1821996	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	793620	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	636169	49.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.56%	
35) Dibromofluoromethane	7.59	113	523449	58.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.62%	
50) Toluene-d8	10.09	98	2442190	50.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.34%	
62) 4-Bromofluorobenzene	12.40	95	842899	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	
Target Compounds						
8) Diethyl Ether	3.41	74	34806	4.687	ug/l	100
16) Acetone	3.82	43	23407	9.248	ug/l	99
18) Methyl Acetate	4.33	43	7628	1.643	ug/l	90
20) Methylene Chloride	4.55	84	906455	65.652	ug/l	97
43) Isopropyl Acetate	8.17	43	57855	1.833	ug/l #	80
95) Naphthalene	15.13	128	343630	11.907	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121918\
Data File : VN053017.D
Acq On : 19 Dec 2018 17:56
Operator : MD\SY
Sample : J6475-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SLUDGE

Quant Time: Dec 20 06:41:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121818W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 18 13:03:37 2018
Response via : Initial Calibration

