

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121919\
 Data File : VN059666.D
 Acq On : 19 Dec 2019 12:47
 Operator : JC/SP
 Sample : VN1219WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1219WBL01

Quant Time: Dec 20 03:15:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	221779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	356117	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	307845	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	108493	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	161266	49.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.94%	
35) Dibromofluoromethane	7.57	113	110993	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
50) Toluene-d8	10.08	98	439687	49.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.68%	
62) 4-Bromofluorobenzene	12.40	95	135627	42.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121919\
Data File : VN059666.D
Acq On : 19 Dec 2019 12:47
Operator : JC/SP
Sample : VN1219WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1219WBL01

Quant Time: Dec 20 03:15:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

