

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121919\
 Data File : VN059669.D
 Acq On : 19 Dec 2019 14:19
 Operator : JC/SP
 Sample : PB125597TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB125597TB

Manual Integrations
 APPROVED

apatel
 12/20/2019 9:20:31 AM

Quant Time: Dec 20 03:25:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	241692	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	390414	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	358908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	135736	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	174099	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
35) Dibromofluoromethane	7.57	113	120219	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
50) Toluene-d8	10.08	98	498526	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
62) 4-Bromofluorobenzene	12.40	95	164747	46.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.36%	
Target Compounds						
20) Methylene Chloride	4.52	84	4007m	1.329	ug/l	
43) Isopropyl Acetate	8.14	43	70322	9.558	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121919\
Data File : VN059669.D
Acq On : 19 Dec 2019 14:19
Operator : JC/SP
Sample : PB125597TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB125597TB

Manual Integrations
APPROVED

apatel
12/20/2019 9:20:31 AM

Quant Time: Dec 20 03:25:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

