

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121919\
 Data File : VN059670.D
 Acq On : 19 Dec 2019 14:41
 Operator : JC/SP
 Sample : K6107-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-06-121819

Quant Time: Dec 20 03:27:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	221374	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	360472	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	334380	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	134502	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	162512	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
35) Dibromofluoromethane	7.57	113	113484	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
50) Toluene-d8	10.08	98	460027	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
62) 4-Bromofluorobenzene	12.40	95	152433	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
Target Compounds						
16) Acetone	3.79	43	9635	5.038	ug/l	95
20) Methylene Chloride	4.52	84	6940	2.513	ug/l #	95
43) Isopropyl Acetate	8.14	43	63168	9.299	ug/l #	90
80) 1,3,5-Trimethylbenzene	12.73	105	9308	1.042	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	24258	2.730	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121919\
Data File : VN059670.D
Acq On : 19 Dec 2019 14:41
Operator : JC/SP
Sample : K6107-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HR-06-121819

Quant Time: Dec 20 03:27:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

