

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121922\
 Data File : VN075875.D
 Acq On : 19 Dec 2022 16:46
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 20 01:30:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	385695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	588954	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	525103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	229006	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	191022	45.525	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.040%
35) Dibromofluoromethane	8.177	113	177237	48.594	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.180%
50) Toluene-d8	10.571	98	691221	50.688	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.380%
62) 4-Bromofluorobenzene	12.853	95	224524	49.710	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121922\
Data File : VN075875.D
Acq On : 19 Dec 2022 16:46
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 20 01:30:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

