

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121922\
 Data File : VN075884.D
 Acq On : 19 Dec 2022 20:22
 Operator : JC\MD
 Sample : N6096-02DL 5X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1DL

Quant Time: Dec 20 01:31:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	416373	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	648026	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	600450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	315307	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	214373	47.326	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.660%		
35) Dibromofluoromethane	8.171	113	187903	46.822	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.640%		
50) Toluene-d8	10.571	98	771210	51.398	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.800%		
62) 4-Bromofluorobenzene	12.853	95	321411	64.674	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	129.340%#		
Target Compounds						
					Qvalue	
31) Cyclohexane	8.230	56	13725	2.116	ug/l #	36
39) Methylcyclohexane	9.600	83	50545	7.859	ug/l	96
52) Toluene	10.629	92	88116	8.226	ug/l	93
67) Ethyl Benzene	11.965	91	408226	19.626	ug/l	98
68) m/p-Xylenes	12.071	106	673996	81.869	ug/l	98
69) o-Xylene	12.400	106	329355	40.426	ug/l	97
73) Isopropylbenzene	12.700	105	282563	12.401	ug/l	100
78) n-propylbenzene	13.041	91	954232	37.970	ug/l	98
80) 1,3,5-Trimethylbenzene	13.176	105	1782159	93.452	ug/l	98
84) 1,2,4-Trimethylbenzene	13.488	105	5729778	302.867	ug/l	99
85) sec-Butylbenzene	13.623	105	744645	32.087	ug/l	82
86) p-Isopropyltoluene	13.735	119	425039	22.011	ug/l	98
89) n-Butylbenzene	14.059	91	656413	41.960	ug/l #	76
95) Naphthalene	15.641	128	257337	16.521	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121922\
Data File : VN075884.D
Acq On : 19 Dec 2022 20:22
Operator : JC\MD
Sample : N6096-02DL 5X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-1DL

Quant Time: Dec 20 01:31:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

