

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080510.D
 Acq On : 19 Dec 2023 10:58
 Operator : JC\MD
 Sample : VN1219MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1219MBL01

Quant Time: Dec 20 03:57:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	192561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	358629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	367322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	164901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	137657	55.971	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 111.940%	
35) Dibromofluoromethane	8.165	113	117184	51.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.080%	
50) Toluene-d8	10.571	98	420974	49.641	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.280%	
62) 4-Bromofluorobenzene	12.847	95	172209	53.478	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080510.D
Acq On : 19 Dec 2023 10:58
Operator : JC\MD
Sample : VN1219MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1219MBL01

Quant Time: Dec 20 03:57:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

