

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080515.D
 Acq On : 19 Dec 2023 13:12
 Operator : JC\MD
 Sample : 05886-07 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1301

Quant Time: Dec 20 04:00:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	235002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	423297	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	439134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	200092	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	162592	54.170	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.340%
35) Dibromofluoromethane	8.165	113	139134	51.841	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.680%
50) Toluene-d8	10.571	98	506955	50.648	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.300%
62) 4-Bromofluorobenzene	12.847	95	200691	52.802	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080515.D
Acq On : 19 Dec 2023 13:12
Operator : JC\MD
Sample : 05886-07 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1301

Quant Time: Dec 20 04:00:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

