

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080516.D
 Acq On : 19 Dec 2023 13:35
 Operator : JC\MD
 Sample : 05886-09 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1302

Quant Time: Dec 20 04:01:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	217513	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	406556	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	423540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193584	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	159073	57.259	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.520%
35) Dibromofluoromethane	8.165	113	136021	52.768	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.540%
50) Toluene-d8	10.565	98	483333	50.276	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.560%
62) 4-Bromofluorobenzene	12.847	95	198706	54.432	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080516.D
Acq On : 19 Dec 2023 13:35
Operator : JC\MD
Sample : 05886-09 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1302

Quant Time: Dec 20 04:01:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

