

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080517.D
 Acq On : 19 Dec 2023 13:59
 Operator : JC\MD
 Sample : 05886-05 200X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1300

Quant Time: Dec 20 04:01:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	219559	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	404039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	417452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159471	56.867	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.740%	
35) Dibromofluoromethane	8.165	113	134609	52.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.100%	
50) Toluene-d8	10.565	98	484750	50.737	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.480%	
62) 4-Bromofluorobenzene	12.853	95	196424	54.142	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.280%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080517.D
Acq On : 19 Dec 2023 13:59
Operator : JC\MD
Sample : 05886-05 200X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1300

Quant Time: Dec 20 04:01:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

