

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080526.D
 Acq On : 19 Dec 2023 17:32
 Operator : JC\MD
 Sample : 05873-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 215-209(TB-07)

Quant Time: Dec 20 04:05:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	214079	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	406608	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	440450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	209027	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	162958	59.598	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 119.200%	
35) Dibromofluoromethane	8.165	113	138267	53.632	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.260%	
50) Toluene-d8	10.565	98	498497	51.847	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.700%	
62) 4-Bromofluorobenzene	12.847	95	207836	56.926	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.860%	
Target Compounds						
16) Acetone	4.430	43	7374	15.210	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080526.D
Acq On : 19 Dec 2023 17:32
Operator : JC\MD
Sample : 05873-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
215-209(TB-07)

Quant Time: Dec 20 04:05:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

