

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080527.D
 Acq On : 19 Dec 2023 17:56
 Operator : JC\MD
 Sample : 05874-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 20 04:05:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	230316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	434602	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	450939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	204904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	165544	56.276	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.560%	
35) Dibromofluoromethane	8.171	113	142392	51.675	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.340%	
50) Toluene-d8	10.565	98	514724	50.086	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.180%	
62) 4-Bromofluorobenzene	12.847	95	213320	54.664	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.320%	
Target Compounds						
16) Acetone	4.430	43	3060	5.867	ug/l	Qvalue # 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080527.D
Acq On : 19 Dec 2023 17:56
Operator : JC\MD
Sample : 05874-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 20 04:05:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

