

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080528.D
 Acq On : 19 Dec 2023 18:20
 Operator : JC\MD
 Sample : 05872-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 215-0207

Quant Time: Dec 20 04:05:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	230804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	429785	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	444591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	195573	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160703	54.515	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.020%	
35) Dibromofluoromethane	8.165	113	137186	50.343	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.680%	
50) Toluene-d8	10.565	98	497457	48.948	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.900%	
62) 4-Bromofluorobenzene	12.853	95	200721	52.012	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.020%	
Target Compounds						Qvalue
16) Acetone	4.430	43	3078	5.889	ug/l	# 77
44) Trichloroethene	9.353	130	7960	2.295	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080528.D
Acq On : 19 Dec 2023 18:20
Operator : JC\MD
Sample : 05872-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
215-0207

Quant Time: Dec 20 04:05:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

