

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080529.D
 Acq On : 19 Dec 2023 18:43
 Operator : JC\MD
 Sample : 05827-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DS01-SO-121323

Quant Time: Dec 20 04:06:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	228561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	423148	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	438130	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	205834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	160617	55.020	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.040%
35) Dibromofluoromethane	8.165	113	137843	51.378	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.760%
50) Toluene-d8	10.565	98	475176	47.489	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.980%
62) 4-Bromofluorobenzene	12.847	95	206863	54.445	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.880%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	80242	155.022	ug/l	# 88
20) Methylene Chloride	5.277	84	13393	4.835	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080529.D
Acq On : 19 Dec 2023 18:43
Operator : JC\MD
Sample : 05827-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DS01-SO-121323

Quant Time: Dec 20 04:06:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

