

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080531.D
 Acq On : 19 Dec 2023 19:31
 Operator : JC\MD
 Sample : 05873-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 215-200

Quant Time: Dec 20 04:07:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	258765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	485381	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	504523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	232706	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	185988	56.274	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 112.540%			
35) Dibromofluoromethane	8.165	113	153893	50.006	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.020%			
50) Toluene-d8	10.565	98	546376	47.604	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.200%			
62) 4-Bromofluorobenzene	12.847	95	236802	54.333	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 108.660%			
Target Compounds					Qvalue	
16) Acetone	4.418	43	41245	70.382	ug/l	97
20) Methylene Chloride	5.271	84	14462	4.611	ug/l	95
43) Isopropyl Acetate	8.688	43	116880	20.630	ug/l #	89
95) Naphthalene	15.641	128	111293	8.580	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080531.D
Acq On : 19 Dec 2023 19:31
Operator : JC\MD
Sample : 05873-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
215-200

Quant Time: Dec 20 04:07:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

