

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
 Data File : VN080533.D
 Acq On : 19 Dec 2023 20:18
 Operator : JC\MD
 Sample : PB157851TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157851TB

Quant Time: Dec 20 04:08:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	241258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	447599	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	459831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	213020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	165706	53.776	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.560%		
35) Dibromofluoromethane	8.165	113	145256	51.183	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.360%		
50) Toluene-d8	10.565	98	528715	49.954	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.900%		
62) 4-Bromofluorobenzene	12.847	95	209811	52.204	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.400%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.506	59	3937	10.849	ug/l #	77
16) Acetone	4.436	43	30713	56.213	ug/l	91
18) Methyl Acetate	5.036	43	3798	1.080	ug/l #	77
20) Methylene Chloride	5.271	84	69741	23.851	ug/l	94
25) 2-Butanone	7.483	43	17301	16.437	ug/l	88
43) Isopropyl Acetate	8.688	43	114028	21.825	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121923\
Data File : VN080533.D
Acq On : 19 Dec 2023 20:18
Operator : JC\MD
Sample : PB157851TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157851TB

Quant Time: Dec 20 04:08:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

