

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
 Data File : VN070235.D
 Acq On : 20 Dec 2021 12:55
 Operator : JC/MD
 Sample : M5044-19 10X
 Misc : 6.60g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DA-19-5.0

Quant Time: Dec 21 03:30:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/21/2021
 Supervised By : Mahesh Dadoda 12/23/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	1116138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2135958	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2556306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1020963	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	866810	50.30	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 100.60%		
35) Dibromofluoromethane	8.021	113	608337	45.47	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 90.94%		
50) Toluene-d8	10.440	98	2886730	53.73	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 107.46%		
62) 4-Bromofluorobenzene	12.731	95	1286136	66.08	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 132.16%		
Target Compounds						
16) Acetone	4.293	43	13680m	1.32	ug/l	Qvalue
84) 1,2,4-Trimethylbenzene	13.366	105	106093	1.52	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
Data File : VN070235.D
Acq On : 20 Dec 2021 12:55
Operator : JC/MD
Sample : M5044-19 10X
Misc : 6.60g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DA-19-5.0

Quant Time: Dec 21 03:30:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/21/2021
Supervised By :Mahesh Dadoda 12/23/2021

