

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
 Data File : VN070238.D
 Acq On : 20 Dec 2021 14:11
 Operator : JC/MD
 Sample : M5106-01
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VAULT-WASTE

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/21/2021
 Supervised By : Mahesh Dadoda 12/23/2021

Quant Time: Dec 20 15:12:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1272424	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	2426446	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2704037	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1452578	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	985801	50.18	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.36%			
35) Dibromofluoromethane	8.016	113	676277	44.50	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 89.00%			
50) Toluene-d8	10.440	98	3268271	53.55	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.10%			
62) 4-Bromofluorobenzene	12.731	95	1508365	68.22	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 136.44%#			
Target Compounds					Qvalue	
16) Acetone	4.304	43	16313m	1.38	ug/l	
25) 2-Butanone	7.348	43	85677	5.34	ug/l	94
39) Methylcyclohexane	9.453	83	135887	4.95	ug/l #	84
73) Isopropylbenzene	12.578	105	708795	5.48	ug/l	98
78) n-propylbenzene	12.919	91	1910034	13.28	ug/l	95
84) 1,2,4-Trimethylbenzene	13.366	105	101575	1.03	ug/l	100
85) sec-Butylbenzene	13.501	105	2212039	17.71	ug/l	92
89) n-Butylbenzene	13.940	91	1501349	18.95	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122021\
Data File : VN070238.D
Acq On : 20 Dec 2021 14:11
Operator : JC/MD
Sample : M5106-01
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VAULT-WASTE

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/21/2021
Supervised By : Mahesh Dadoda 12/23/2021

Quant Time: Dec 20 15:12:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

